

The NJDA Plant Laboratory is seeking a **Leica S9D microscope and camera with internal WiFi capability**. Currently the NJDA Plant Laboratory does not possess a microscope of adequate quality to be able to perform necessary lab screening for the Medical Marijuana and Nursery Programs. The specifications required include compatibility with high grade cameras (needed to document NJDA lab work and produce images for publication online and elsewhere), at least 55x zoom, and a generous amount of space between the specimen and the lenses (needed to perform dissections efficiently.)

STATE OF NEW JERSEY
DEPARTMENT OF THE TREASURY
AGENCY REQUEST FOR PROPOSAL

VENDOR NAME AND ADDRESS:		RETURN THIS PROPOSAL TO:		DELIVER TO:	
SBE CATEGORY:		FAX NUMBER:			
NOTE: This proposal must be received by the opening date/time at the place named above.		AGENCY PERSON TO CONTACT:			
FISCAL YEAR:	ACCOUNT NUMBER:	AGENCY REFERENCE NUMBER:		COMMODITY CODE NUMBER:	
ITEM NUMBER	QUANTITY	UNIT	DESCRIPTION (ALL ITEMS MUST BE DELIVERED F.O.B. DESTINATION)	UNIT PRICE	AMOUNT
PRICES ARE FIRM UNTIL THE FOLLOWING DATE:				TOTAL	
CASH DISCOUNT:	DATE OF DELIVERY:	VENDOR'S FEDERAL I.D. NUMBER:		VENDOR'S TELEPHONE NUMBER:	
VENDOR'S SIGNATURE (Must be Signed):		PRINT OR TYPE NAME BELOW:		DATE:	