

# RETURN & REPAIR FORM



CUSTOMER Township of Edison Cust. # \_\_\_\_\_  
 Date ret. 12/6 Carrier Fed Ex Method Priority RMA# \_\_\_\_\_  
 7110 Serial# AKL0-0012 Warranty exp.: \_\_\_\_\_

Reported Problem: fuel cell replacement

- Accessories (check all that apply):  Regulator  P.Paper  Scotty V  Printer Rib.  Pouch  
 Sim. Temp. Probe Ser.# \_\_\_\_\_  Keyboard  Organizer  Key  
 Mouthpieces \_\_\_\_\_  Sim. To cuvette hose  Pump to Sim. hose  Power cord  
 Other (specify) \_\_\_\_\_  
 Other (specify) \_\_\_\_\_

<u>Part#</u>	<u>Description</u>	<u>\$</u>	<u>Qty</u>	<u>Total</u>
<u>6808455</u>	<u>Fuel Cell</u>		<u>1</u>	
<u>MPCAL 7H</u>			<u>1</u>	
		Repair Time	<u>.5</u>	Time \$

TOTAL REPAIR \$ \_\_\_\_\_

REPAIR NOTES:  
Replaced fuel cell. Upgraded Unit to NS-3.11.  
Calibrated Unit, full QC & OPS CHK.

Date 08 DEC 04 Technician [Signature]  
 Every above checked item is included  Shipping \_\_\_\_\_