

Return & Repair Form:

Customer Information:

Company Name: New Jersey State Police B: _____ S: _____
 Date Received: 1/16/09 Date Given to Service: _____
 Carrier: FedEx - (UPS) - DHL - USPS Method: (GRD) - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 (7110) - 6510 - 6810 Serial #: AR L121 0280
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:


Reason for Return: LOW IR RESULTS Test # _____

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FUEL CELL</u>	<u>1</u>	
<u>MPLR71</u>	<u>CAL</u>	<u>1</u>	
<u>MPLABOR</u>		<u>5</u>	

Repair Notes:

FC IS VF = JUNE 2004 / REPLACED / HAS LIE 303 DETECTOR /
IR FILTER LOOSE - SECURED IN PLACE / SPOTS ON MIRRORS -
BLEW OUT W/ AIR.

CAL W/ QC & OPS CHECKS

Service Technician: 

Date: 01-20-09