RETURN & REPAIR FORM

ALCOTEST 7110

CUSTOMER NJSP Cust. #

Date ret. 7-15 Carrier FED EX Method PRIORITY RMA#

7110 Serial# ARLM-0282 Warranty exp.: WARRANTY

Reported Problem: ERROR 41

Accessories (check all that apply): ☐ Regulator ☐ P.Paper ☐ Scotty V ☐ PrinterRib. ☐ Pouch
Sim.Temp.Probe Ser.# ☐ Keyboard ☐ Organizer ☐ Key
☐ Mouthpieces ☐ Sim. to cuvette hose ☐ Pump to Sim. hose ☐ Power cord
☐ Other (specify)
☐ Other (specify)

Part#	Description	$	Qty	Total
W 8100902	FLOW SENSOR		1	
W 8100809	BREATH HOSE		1	
		Repair Time	1.0	Time $ N/C

TOTAL REPAIR $

REPAIR NOTES:
E-41 - FLOW SENSOR BROKEN/REPLACED/FLOW CAL
QC CHECK - RESULTS SLIGHTLY HIGH
CAL-TEST-FULL QC CHECK
CHECK NJ SPEC CONFIG.
REPLACE - BREATH HOSE - CAL AND RETEST ACC
VIA HOSE

Date 07-16-98 Technician
Every above checked item is included ☐ Shipping

G:\HOME\COMMON\FORMS\RRFORM.WPD