

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Don Pouliot NJST Cust # _____
 Date of Return: 6/19/06 Carrier: Fed ex Method: Express Saver

Product Information:

Product: 7110 7410 6510 6810 Serial # ARLM-0285
 Description: A B PLUS SCREENER _____ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter
- Regulator
- Mag Card Rdr # _____
- Printer Paper
- Printer Ribbon
- Casio # _____
- Mouthpieces
- Carrying Case
- Dry Gas
- Other (specify): Power Supply

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: ERROR 009

Part #	Description	Qty	Total
12069	Power Supply	1	
MPCAL71	7110 CAL	1	

Repair Notes: Replaced power supply due to error 009.
Cal ops check + O.C.

Technician: CRD

Date: 6/20/06