

Return & Repair Form:

Customer Information:

Company Name: NOSP/DOO B: _____ S: _____
 Date Received: 11-19-08 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P - _____
 Serial #: AR LM - 0285
 Printer Ser #: AR - _____
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: FL

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FL</u>	<u>1</u>	<u>w/c warr</u>
<u>MP CAL 71</u>	<u>cal</u>	<u>1</u>	<u>w/c warr</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>w/c warr</u>

Repair Notes:

FL ARUF-0679 replaced with FL ARHJ-0283

cal w/DC & ops check

Service Technician: [Signature]

Date: 12/01/09