

Return & Repair Form

Customer Information:

Company Name: New Jersey State PD B: _____ S: _____
 Date Received: 7-18-08 Date Given to Service: _____
 Carrier: FedEx UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 7110 - 6510 - 6810 Serial #: AR 617 - 0286
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: ERROR 43

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>192⁰⁰</u>
<u>MPCAL 71</u>	<u>CAL</u>	<u>1</u>	<u>120⁰⁰</u>
<u>MPCAL BOR</u>		<u>.5</u>	<u>44⁰⁰</u>

Repair Notes:

ERROR 43 NOT DUPLICATED

↳ +S/H/I

REPLACE FC / CAL w/ QC EOPS CHECKS

Service Technician: [Signature]

Date: 07-29-08