

RETURN & REPAIR FORM



CUSTOMER NJ State Police Cust. # _____

Date ret. 5/25 Carrier UPS Method G1D RMA# _____

7110 Serial# ARLM 0288 Warranty exp.: _____

Reported Problem: _____

- Accessories (check all that apply): Regulator P.Paper Scotty V Printer Rib. Pouch
 Sim. Temp. Probe Ser.# _____ Keyboard Organizer Key
 Mouthpieces _____ Sim. To cuvette hose Pump to Sim. hose Power cord
 Other (specify) _____
 Other (specify) _____

<u>Part#</u>	<u>Description</u>	<u>\$</u>	<u>Qty</u>	<u>Total</u>
w/ 6809485	Fuel Cell		1	
w/ 12114	Screw		1	
w/ 6809514	Breath-hose		1	
w/ 12117	Screw		1	
w/ 6809480	HOSE COVER		1	
w/ 12118	WASHER		1	
MPCAL 71			1	
	Repair Time		.5	Time \$

TOTAL REPAIR \$ _____

REPAIR NOTES:
 Replaced fuel cell. Replaced Breath-hose. Upgraded
 Firmware TO NS-3.10. Full QC & OPS CHK WITH
 CALIBRATION.

Date 26 Oct 04 Technician EW
 Total price of item (incl. tax) _____ Shipping _____