

## Return & Repair Form

### Customer Information:

CUSTOMER NJSP Cust # \_\_\_\_\_  
 Date of Return: 1/30/07 Carrier: UPS Method: 2 Day

### Product Information:

Product : 7110 7410 6510 6810 Serial # ARLM-0288  
 Description: A B PLUS SCREENER Printer Serial # \_\_\_\_\_  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER : \_\_\_\_\_ Probe Serial # \_\_\_\_\_  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
**Accessories:** (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: Will not power up

Part #	Description	Qty	Total
MPCALTI	7110 CAL	1	
MPLABOR	LABOR	.5	

Repair Notes: Electrical connection had come loose on switch circuitry. Reconnected and not glued connection. Full cal ops check + QC.

Technician: CRD

Date: 1/31/07