

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER HANOVER Police Dept. Cust # _____
 Date of Return: 3/30/17 Carrier: FED EX Method: Priority

Product Information:

Product : 7110 7410 6510 6810 Serial # ARNK-0021
Description: A B PLUS SCREENER Printer Serial # _____
Whole Inst. Top Half Simulator Serial # _____
OTHER : _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: repair per attached EC Problem

Part #	Description	Qty	Total
6808455	Fuel Cell	1	
MPLABOR	LABOR	.5	
MPCAL71	7110 CAL	1	

Repair Notes: Replaced leaky F/C. Full cal ops check + O.C.

Technician: CRO

Date: 4/2/07