

Return & Repair Form

Customer Information:

CUSTOMER NJSP Cust # 53075/53076  
 Date of Return: 10/2/2007 Carrier: FX UPS DHL USPS Method: GRD NDA 2DY 3DY  
 OTHER: \_\_\_\_\_

Product Information:

Product : 7110 7410 6510 6810 Serial # ARAK - 0036  
 Description: A B PLUS SCREENER \_ Printer Serial # AR -  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER : \_\_\_\_\_ Probe Serial # DD P -  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
 Accessories: (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

Repair Information:

Test # \_\_\_\_\_

Reason for Return: \_\_\_\_\_

Part #	Description	Qty	Total
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>WARR</u>
<u>MPCAL 71</u>	<u>7110 CMC</u>	<u>1</u>	<u>WARR</u>
<u>6809512</u>	<u>HOSE BARB</u>	<u>1</u>	<u>WARR</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>WARR</u>

Repair Notes: REPLACE FC - REPLACE HOSE BARB - CAL QC OPS CHECKS

Technician: [Signature]

Date: 10-03-07