

# RETURN & REPAIR FORM



CUSTOMER S. Plainfield police Cust. # \_\_\_\_\_

Date ret. 7/30 Carrier FedEx Method SWEN RMA# \_\_\_\_\_

7110 Serial# FRNK-0056 Warranty exp.: \_\_\_\_\_

Reported Problem: \_\_\_\_\_

- Accessories (check all that apply):  Regulator  P.Paper  Scotty V  Printer Rib.  Pouch  
 Sim. Temp. Probe Ser.# \_\_\_\_\_  Keyboard  Organizer  Key  
 Mouthpieces \_\_\_\_\_  Sim. To cuvette hose  Pump to Sim. hose  Power cord  
 Other (specify) \_\_\_\_\_  
 Other (specify) \_\_\_\_\_

Part#	Description	\$	Qty	Total
6808455 ARUF-0728	Fuel Cell		1	
MPCAL 71		Repair Time	1 .5	Time \$

TOTAL REPAIR \$ \_\_\_\_\_

REPAIR NOTES:  
Replaced Fuel Cell. Calibrated Unit Full DC  
& OPS CHKS. UPGRADED UNIT TO NJ-3.8.

Date 2 Aug 04 Technician [Signature]  
 Every above checked item is included  Shipping