

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Delaware River Port Authority P.D., Camden, NJ

Date Received: 04/29/15

Date given to service: 04/29/15

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☐ 3DAY ☐ 2DAY
☒ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR RL-0048

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC
4414166	Labor	.5	NC
6808455	Fuel Cell	1	NC

Repair Notes: Replaced defective fuel cell.

Fuel Cell = ARHA-1574

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 05/11/15