

Return & Repair Form:

Customer Information:

Company Name: Morris Plains Boro B: _____ S: _____

Date Received: 6-26-09 Date Given to Service: _____

Carrier: FedEx UPS - DHL - USPS Method: GRD - NDA - 2DY 3DY Other _____

Product Information:

Product: 7410 7110 6510 - 6810

Serial #: AR SA - 0014

Description: A - B - Plus - Screener - Demo

Printer Ser #: AR -

Whole Instrument

Top ½

Sim Ser #: _____

Other _____

Probe Ser # DD _____ P -

ACCESSORIES:

☐ 110 V A/C Adapter

☐ Regulator

☐ Mag Card Rdr# _____

☐ Printer Paper

☐ Printer Ribbon

☐ Casio # _____

☐ Mouthpieces

☐ Carrying Case

☐ Dry Gas

Other (Please Specify) _____

Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: ERROR 32 - FC + 1R?

Part Number	Description	Qty	Total Cost
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C WARR</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>6809500</u>	<u>Cuvette</u>	<u>1</u>	<u>N/C WARR</u>
<u>6809512</u>	<u>Grommit with Ventili</u>	<u>1</u>	<u>N/C WARR</u>
<u>MA Labor</u>	<u>labor</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes:

Replace Cuvette, FC, Grommit with Ventili
Cal w/Qc ; ops check

Service Technician: JS

Date: 06/26/2009