

# RETURN AND REPAIR FORM

# Dräger

## Customer Information

B:

S:

Company Name: Stafford Twp PD , Manahawkin , NJ

Date Received: 06/29/12

Date given to service: 06/29/12

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY  
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810  
☒ 7110 ☐ 8610  
☐ 7510 ☐ DT5000  
☐ 7410 Upper-half ☐ 9510  
☐ 7410 Whole

Serial Number: AR SA-0031

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

## Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper  
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case  
☐ Dry Gas ☐ Other (please specify) \_\_\_\_\_

## Repair Information

Test#:

| Part Number | Description | Qty | Total Cost |
|-------------|-------------|-----|------------|
|             |             |     |            |
|             |             |     |            |
| MP Cal 71   | Calibration | 1   | NC/W       |
| 6808455     | Fuel cell   | 1   | NC/W       |
|             |             |     |            |
| MP Labor    | Labor       | .5  | NC/W       |

Repair Notes: Replaced defective fuel cell with fuel cell ARDD-0625.

CAL W/QC AND OPS CHECK

Service Technician JS

Date: 07-09-12