

Return & Repair Form:

Customer Information:

Company Name: Highland Park B: _____ S: _____

Date Received: 4-12-10 Date Given to Service: _____

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY Other _____

Product Information:

Product: 7410-7110-6510-6810-7510
9510- Drug Tester 5000

Serial #: AR SC - 0007

Description: A - B - Plus - Screener - Demo

Printer Ser #: AR -

Whole Instrument

Top ½

Sim Ser #: _____

Other _____

Probe Ser # DD P -

ACCESSORIES:

☐ 110 V A/C Adapter

☐ Regulator

☐ Mag Card Rdr# _____

☐ Printer Paper

☐ Printer Ribbon

☐ Casio # _____

☐ Mouthpieces

☐ Carrying Case

☐ Dry Gas

Other (Please Specify) _____

Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:

FC

Part Number	Description	Qty	Total Cost
<u>Q808455</u>	<u>FC</u>	<u>1</u>	<u>N/C W</u>
<u>Q808486</u>	<u>Plate</u>	<u>2</u>	<u>N/C W</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C W</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

Repair Notes:

NEW FC HRAW - 0985

cal w/QC! Ops check

Service Technician: _____

Date: 04/12/2010