

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER DENVILLE PD NJ Cust # 51560
 Date of Return: 12/1 Carrier: FedEx Method: PRIORITY

Product Information:

Product: 7110 7410 6510 6810 Serial # ARSC-0010
 Description: A B PLUS SCREENER _ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____
 Demo Unit ☐ Returned to stock on: _____
 Accessories: (check all that apply):
☐ 110VAC Adapter ☐ Regulator ☐ Mag Card Rdr # _____
☐ Printer Paper ☐ Printer Ribbon ☐ Casio # _____
☐ Mouthpieces ☐ Carrying Case ☐ Dry Gas
☐ Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Error (032) Pump, Motor

Part #	Description	Qty	Total
<u>6810269</u>	<u>EC Motor</u>	<u>1</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Replaced EC Motor because it was seized. Cal ops check & Q.C.

Technician: CRDDate: 12/1/05