

Return & Repair Form

Customer Information:

CUSTOMER SGT KAZ DENVILLE TWP RD Cust # 15-51560
 Date of Return: 7-12 Carrier: FX Method: PRI-STD

Product Information:

Product: 7110 7410 6510 6810 Serial # ARSC-0010
 Description: A B PLUS SCREENER _ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____

Demo Unit ☐ Returned to stock on: _____

Accessories: (check all that apply):

- ☐ 110VAC Adapter ☐ Regulator ☐ Mag Card Rdr # _____
☐ Printer Paper ☐ Printer Ribbon ☐ Casio # _____
☐ Mouthpieces ☐ Carrying Case ☐ Dry Gas
 Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: FUEL CELL PROBLEM?

Part #	Description	Qty	Total
6808455	FUEL CELL	1	WARRANTY
MPCAL71	7110 CAL	1	N/C
MPLABOR	LABOR	.5	N/C

Repair Notes: Replace FUEL CELL w/ARYA-0431
CAL w/ OPS CHECKS

BASED ON DATA PROVIDED - WE REPLACED THE FC

Technician: IMB/AC

Date: 07.13.07