

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Denville Police, Denville, NJ

Date Received: 02/12/14

Date given to service: 02/12/14

Carrier: ☐ FedEx ☒ UPS ☐ USPS

Shipping Method: ☒ GRD ☐ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR SC-0010

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: April 2016

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
MP Labor	Labor	1	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W
MP Misc.	Keyboard Port	1	NC/W

Repair Notes: Replaced fuel cell, spacer plates, and keyboard port..

Keyboard port was damaged.

Fuel Cell = AREL-1276

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 02/20/14