

Return & Repair Form

Customer Information:

CUSTOMER DELAWARE RIVER N.J. Cust # 51616/51618
 Date of Return: 10/16/2007 Carrier: FX UPS DHL USPS Method: GRD NDA 2DY 3DY
 OTHER: STP

Product Information:

Product : 7110 7410 6510 6810 Serial # AR SC - 0027
 Description: A B PLUS SCREENER _ Printer Serial # AR -
 Whole Inst. Top Half Simulator Serial # _____
 OTHER : _____ Probe Serial # DD P -
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: FC

Part #	Description	Qty	Total
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>WARR</u>
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	<u>WARR</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>WARR</u>

Repair Notes: REPLACE CELL - CAL QC & OPS CHECKS.

Technician: [Signature]

Date: 10.16.07