

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Mt. Laurel Twp. P.D., Mt. Laurel, NJ

Date Received: 10/19/2016

Date given to service: 10/19/2016

Carrier: ☒ FedEx

☐ UPS

☐ USPS

Shipping Method:

☐ GRD

☒ 3DAY

☐ 2DAY

☐ NDA-PRI

☐ NDA-STD

Product:

☐ 6510

☐ 6810

☒ 7110

☐ 8610

Serial Number: AR SC-0047

☐ 7510

☐ DT5000

Printer Serial#: AR

☐ 7410 Upper-half

☐ 9510

Sim Serial#: DD

☐ 7410 Whole

Probe Serial#: DD

Warranty Expires: December 2016

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter

☐ Regulator

☐ Printer Ribbon

☐ Printer Paper

☐ Mouthpieces

☐ 9510 Stylus

☐ 9510 Top Cover

☐ Carrying Case

☐ Dry Gas

☐ Other (please specify) _____

Repair Information

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	3	NC/W
Misc.	Power Supply (Refurbished)	1	NC

Repair Notes: Replaced defective power supply (refurbished).

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 11/01/2016