

Return & Repair Form

Customer Information:

CUSTOMER Plainsboro Twp. Cust # _____
 Date of Return: 1/22/07 Carrier: FedEx Method: 2 Day

Product Information:

Product: 7110 7410 6510 6810 Serial # ARTL-0005
 Description: A B PLUS SCREENER Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____
 Demo Unit ☐ Returned to stock on: _____
 Accessories: (check all that apply):
☐ 110VAC Adapter ☐ Regulator ☐ Mag Card Rdr # _____
☐ Printer Paper ☐ Printer Ribbon ☐ Casio # _____
☐ Mouthpieces ☐ Carrying Case ☐ Dry Gas
☐ Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: EC Results drop on SOLN-CHANGE

Part #	Description	Qty	Total
6808455	Fuel Cell	1	
MPLA 671	7110 CAL	1	
MPLABOR	LABOR	.5	

Repair Notes: Replaced leaky F/C. Full cal ops
check + O.C.

Technician: CRO

Date: 1/22/07