

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: South River Boro P.D., South River, NJ

Date Received: 08/07/2017

Date given to service: 08/07/2017

Carrier: ☐ FedEx

☐ UPS

☐ USPS

Shipping Method:

☐ GRD

☐ NDA-PRI

☐ 3DAY

☐ NDA-STD

☐ 2DAY

Product: ☐ 6510

☐ 6810

☒ 7110

☐ 8610

Serial Number: AR TL-0006

☐ 7510

☐ DT5000

Printer Serial#: AR

☐ 7410 Upper-half

☐ 9510

Sim Serial#: DD

☐ 7410 Whole

Probe Serial#: DD

Warranty Expires: _____

Description:

☐ A

☐ B

☐ Plus

☐ Demo

☐ Screener

☐ Trade In

Accessories

☐ 110V A/C Adapter

☐ Regulator

☐ Printer Ribbon

☐ Printer Paper

☐ Mouthpieces

☐ 9510 Stylus

☐ 9510 Top Cover

☐ Carrying Case

☐ Dry Gas

☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	.5	
6808455	Fuel Cell	1	
6808486	Spacer Plates	2	

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = ARJE-1840

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 08/31/17