

609-845-5300
1011 Cooper St.
Deptford NJ 08096

Drägersafety

Internal Use Only

Return & Repair Form

Customer Information:

CUSTOMER Deptford Township Cust # 54540
Date of Return: 11/2/06 Carrier: Feed Ex Method: Priority

Product Information:

Product: (7110) 7410 6510 6810 Serial # ARTL-0007

Description: A B PLUS SCREENER _ Printer Serial # _____

Whole Inst. Top Half Simulator Serial # _____

OTHER: _____ Probe Serial # _____

Demo Unit ☐ Returned to stock on: _____

Accessories: (check all that apply):

- | | | |
|---|---|---|
| ☐ 110VAC Adapter | ☐ Regulator | ☐ Mag Card Rdr # _____ |
| ☐ Printer Paper | ☐ Printer Ribbon | ☐ Casio # _____ |
| ☐ Mouthpieces | ☐ Carrying Case | ☐ Dry Gas |
| ☐ Other (specify): _____ | | |

Warranty Expiration Date: _____

Repair Information: Test # _____

Reason for Return: ~~IR + EC erratic~~ IR + EC erratic.

Part #	Description	Qty	Total
6808455	Fuel Cell	1	
MPLABOR	LABOR	1	
6809512	Cuvette port w/ check valve	1	
MPCAL71	7110 CAL	1	

Repair Notes: Replaced F/C due to possible leak. Replaced
check valve in cuvette to improve stability.
Cal ops check + QC.

Technician: CRD

Date: 11/7/06