

# RETURN AND REPAIR FORM

# Dräger

## Customer Information

B:

S:

Company Name: Deptford Twp P.D., Deptford N.J.

Date Received: 08/09/11 Date given to service: \_\_\_\_\_

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY  
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810  
☒ 7110 ☐ 8610  
☐ 7510 ☐ DT5000  
☐ 7410 Upper-half ☐ 9510  
☐ 7410 Whole

Serial Number: AR TL - 0007

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

## Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper  
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case  
☐ Dry Gas ☐ Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

| Part Number      | Description      | Qty       | Total Cost   |
|------------------|------------------|-----------|--------------|
| <u>6808455</u>   | <u>Fuel Cell</u> | <u>1</u>  | <u>n/c w</u> |
| <u>mp cal 21</u> | <u>Calibrate</u> | <u>1</u>  | <u>n/c w</u> |
|                  |                  |           |              |
|                  |                  |           |              |
|                  |                  |           |              |
| <u>mp labor</u>  | <u>labor</u>     | <u>.5</u> | <u>n/c w</u> |

Repair Notes: NEW FUEL CELL ARCA-0910

Cal/w @ C & Ops Check

Service Technician: [Signature] Date: 8-10-11