

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: East Brunswick Twp. Police Dept., East Brunswick, NJ

Date Received: 01/28/14

Date given to service: 01/28/14

Carrier: ☒ FedEx ☐ UPS ☐ USPS

Shipping Method: ☐ GRD ☐ NDA-PRI

☒ 3DAY ☐ 2DAY
☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR.TL-0008

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter
☐ Mouthpieces
☐ Dry Gas

☐ Regulator
☐ 9510 Stylus
☐ Other (please specify) _____

☐ Printer Ribbon
☐ 9510 Top Cover

☐ Printer Paper
☐ Carrying Case

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
MP Labor	Labor	10	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W
8315075	Alcotest 7110 Motherboard	1	NC/W
6808374	Geared Engine (Motor)	1	NC/W
6809514	Breath Hose	1	NC/W

Repair Notes: Replaced, fuel cell, spacer plates, breath hose, motherboard, and geared engine.

Motherboard was replaced due to unit not creating file numbers correctly and causing No Admittant errors. Breath hose was changed due to BLOWING NOT ALLOWED issues. Geared Engine was changed due to issue with it getting stuck.

Fuel Cell = AREL-1308

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 02/13/14