

Dräger safety

Return & Repair Form**Customer Information:**

CUSTOMER South Brunswick Twp. P.D. Cust # _____
 Date of Return: 3/30/7 Carrier: Fed Ex Method: Exp. Saver

Product Information:

Product : 7110 7410 6510 6810 Serial # ARTL-0014
 Description: A B PLUS SCREENER Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER : _____ Probe Serial # _____

Demo Unit ☐ Returned to stock on: _____

Accessories: (check all that apply):

- ☐ 110VAC Adapter ☐ Regulator ☐ Mag Card Rdr # _____
☐ Printer Paper ☐ Printer Ribbon ☐ Casio # _____
☐ Mouthpieces ☐ Carrying Case ☐ Dry Gas
☒ Other (specify): Power CARD

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Repair per attached EC problem

Part #	Description	Qty	Total
6808455	Fuel Cell	1	
MPLAB02	LABOR	.5	
MPCAL71	7110 CAL	1	

Repair Notes: leaky.
Replaced questionable F/C. Full cal
ops check + Q.C.

Technician: ced

Date: 4/2/07