

Dräger safety

Return & Repair Form**Customer Information:**

CUSTOMER Stuyvesant PD Cust # 53883
 Date of Return: 1/12 Carrier: Fed Ex Method: Priority

Product Information:

Product: 110 7410 6510 6810 Serial # ARTZ-0015
 Description: A B PLUS SCREENER _ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____

Demo Unit ☐ Returned to stock on: _____

Accessories: (check all that apply):

- ☐ 110VAC Adapter ☐ Regulator ☐ Mag Card Rdr # _____
☐ Printer Paper ☐ Printer Ribbon ☐ Casio # _____
☐ Mouthpieces ☐ Carrying Case ☐ Dry Gas
☐ Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: See AD High EC Readings

Part #	Description	Qty	Total
0808455	Fuel Cell	1	
MP CAL 71	7110 CAL	1	
MP LABOR	LABOR	.5	

Repair Notes: Replaced Fuel cell due to leakage. Cal
ops check - O.C.

Technician: CRD

Date: 1/12/04