

# Return & Repair Form:

## Customer Information:

Company Name: Morris Twp Police B: \_\_\_\_\_  
Date Received: 04/11/2010 Date Given to Service: \_\_\_\_\_  
Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-7110-6510-6810-7510  
9510- Drug Tester 5000  
Description: A - B - Plus - Screener - Demo  
Whole Instrument Top 1/2  
Other \_\_\_\_\_  
Serial #: AR TL - 0017  
Printer Ser #: AR \_\_\_\_\_  
Sim Ser #: \_\_\_\_\_  
Probe Ser # DD \_\_\_\_\_ P \_\_\_\_\_  
**ACCESSORIES:**  
☐ 110 V A/C Adapter ☐ Regulator ☐ Mag Card Rdr# \_\_\_\_\_  
☐ Printer Paper ☐ Printer Ribbon ☐ Casio # \_\_\_\_\_  
☐ Mouthpieces ☐ Carrying Case ☐ Dry Gas  
Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return:

FC

| Part Number      | Description  | Qty       | Total Cost      |
|------------------|--------------|-----------|-----------------|
| <u>6808455</u>   | <u>FC</u>    | <u>1</u>  | <u>N/C Warr</u> |
| <u>6808486</u>   | <u>Plate</u> | <u>1</u>  | <u>N/C Warr</u> |
| <u>MP Cal 71</u> | <u>Cal</u>   | <u>1</u>  | <u>N/C Warr</u> |
| <u>MP Labor</u>  | <u>Labor</u> | <u>.5</u> | <u>N/C Warr</u> |

Repair Notes:

NEW FC # ARAN-1031

Cal w/80 / 80ps check

Service Technician: JL

Date: 4/16/2010