

Return & Repair Form

Customer Information:

CUSTOMER City of Brigantine Cust # 53947
 Date of Return: 9/25/2007 Carrier: FX UPS DHL USPS Method: GRD NDA 2DY 3DY
 OTHER: priority ON

Product Information:

Product : 7110 7410 6510 6810 Serial # AR TL - 0019
 Description: A B PLUS SCREENER _ Printer Serial # AR -
 Whole Inst. Top Half Simulator Serial # _____
 OTHER : _____ Probe Serial # DD P -
 Demo Unit ☐ Returned to stock on: _____
 Accessories: (check all that apply):
☐ 110VAC Adapter ☐ Regulator ☐ Mag Card Rdr # _____
☐ Printer Paper ☐ Printer Ribbon ☐ Casio # _____
☐ Mouthpieces ☐ Carrying Case ☐ Dry Gas
☐ Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: FUEL CELL PROBLEM

Part #	Description	Qty	Total
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>WARR</u>
<u>MPCAL 71</u>	<u>7110 CAL</u>	<u>1</u>	<u>WARR</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>WARR</u>

Repair Notes: REPLACE CELL / CAL, QC, OPS CHECKS

Technician: [Signature]

Date: 9-27-07