

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Brigantine City P.D., Brigantine N.J.

Date Received: 10/21/11 Date given to service: _____

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY
☐ NDA-PRI ☒ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole
 Serial Number: AR TL-0019
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
6807819	micro switch	1	w/c w
6808455	Fuel Cell	1	w/c w
mp cal 71	calibration	1	w/c w

Repair Notes: Replace Fuel Cell ARAE-0877, because of low
EL reading, replaced with Fuel Cell ARCH-0444.

Also replaced the micro switch, it was damaged
during 7110 inspection.

cal w/DC; ops check

Service Technician: [Signature] Date: 10/24/11