

Internal Use Only

Customer Information:

Cust #

Method: ~~PRIORITY~~ STD

Serial # ARTL-0023

Printer Serial #

Simulator Serial #

Probe Serial #

Warranty Expiration Date:

Test #Reason for Return: LOW IR/EC RESULTS

Part #	Description	Qty	Total
MPLABOR	LABOR	2 HR	N/C

Repair Notes: PERFORMED 25+25 STAND CHECK W/ SINGLE SIM & NS
SOLUTION LOT # 06K037 69/1000 / NOTED TYPICAL DEPLETION OF
SOLUTION CONCENTRATION AS EXPECTED. REMOVED - DISASSEMBLED -
AND BLOW OUT CUVETTE W/ AIR - RE-ASSEMBLED - MULTIPLE ACC
TESTS PERFORMED - NO CAL REQUIRED - NO CHANGE
TO CAL MADE

Technician:

Date: 07-20-07