

Return & Repair Form:

Customer Information:

Company Name: Barlington TWP B: _____ S: _____
 Date Received: 3-1-10 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-~~7110~~6510-6810-7510 Serial #: AR TH -0024
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>mp cal 71</u>	<u>cal</u>	<u>1</u>	<u>N/C WARR</u>
<u>8312074</u>	<u>Intake Tube</u>	<u>1</u>	<u>N/C WARR</u>
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>N/C WARR</u>
<u>mp labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes:

New FC # ARAN-1066
Place Intake Tube because of wear.
Cal w/QC & ops check

Service Technician: JS

Date: 03/04/2010