

1-30-07
PER SGT DENTON
RETURN UNIT VIA FY PRI.
@

Internal Use Only

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Carteret Boro P.D. Cust # _____
Date of Return: 1/29/07 Carrier: FedEx Method: Ground

Product Information:

Product: 7110 7410 6510 6810 Serial # ARTL-0026

Description: A B PLUS SCREENER Printer Serial # _____

Whole Inst. Top Half Simulator Serial # _____

OTHER: _____ Probe Serial # _____

Demo Unit ☐ Returned to stock on: _____

Accessories: (check all that apply):

- | | | |
|---|---|---|
| ☐ 110VAC Adapter | ☐ Regulator | ☐ Mag Card Rdr # _____ |
| ☐ Printer Paper | ☐ Printer Ribbon | ☐ Casio # _____ |
| ☐ Mouthpieces | ☐ Carrying Case | ☐ Dry Gas |
| ☐ Other (specify): _____ | | |

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Unit Randomly shuts off / Low EC readings

Part #	Description	Qty	Total
6808455	Fuel Cell	1	
34901	Power Cord	1	
MPLABOR	LABOR	5	
MPCAL71	7110 CAL	1	

Repair Notes: Replaced leaky Fuel Cell and faulty power cord. Full cal, ops check + Q.C.

Technician: CRD

Date: 1/30/07