

Dräger safety

Return & Repair Form**Customer Information:**

CUSTOMER Carteret Boro PD Cust # _____
 Date of Return: 3/26/07 Carrier: FedEx Method: Express

Product Information:

Product: 7110 7410 6510 6810 Serial # APIL-0024
 Description: A B PLUS SCREENER Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____
 Demo Unit ☐ Returned to stock on: _____ Power Cord.

Accessories: (check all that apply):

- ☐ 110VAC Adapter ☐ Regulator ☐ Mag Card Rdr # _____
☐ Printer Paper ☐ Printer Ribbon ☐ Casio # _____
☐ Mouthpieces ☐ Carrying Case ☐ Dry Gas
☐ Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: FL Problem

Part #	Description	Qty	Total
<u>6308455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Replaced questionable FLC - Full cal
ops check + Q.C.

Technician: CRDDate: 3/26/07