

Dräger safety

Internal Use Only

Return & Repair Form

Customer Information:

CUSTOMER Monroe PD Cust # 82226  
Date of Return: 3/28 Carrier: FedEx Method: PRIORITY

Product Information:

Product: 7110 7410 6510 6810 Serial # ARTL-0030  
Description: A B PLUS SCREENER \_ Printer Serial # \_\_\_\_\_  
Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
OTHER: \_\_\_\_\_ Probe Serial # \_\_\_\_\_

Demo Unit ☐ Returned to stock on: \_\_\_\_\_

Accessories: (check all that apply):

- ☐ 110VAC Adapter ☐ Regulator ☐ Mag Card Rdr # \_\_\_\_\_  
☐ Printer Paper ☐ Printer Ribbon ☐ Casio # \_\_\_\_\_  
☐ Mouthpieces ☐ Carrying Case ☐ Dry Gas \_\_\_\_\_  
☐ Other (specify): \_\_\_\_\_

Warranty Expiration Date: \_\_\_\_\_

Repair Information:

Test # \_\_\_\_\_

Reason for Return: Missing Data on printouts

| Part #  | Description  | Qty | Total |
|---------|--------------|-----|-------|
| 8315075 | Mother Board | 1   |       |
| MPCAL71 | 7110 CAL     | 1   |       |
|         |              |     |       |
|         |              |     |       |
|         |              |     |       |
|         |              |     |       |

Repair Notes: Could not reproduce problem of missing data.  
Replaced mother board per engineering. Cal ops check  
and QC.

Technician: CRD

Date: 4/18/06