

Return & Repair Form

Customer Information:

Company Name: MONROE TWP POLICE B: _____ S: _____

Date Received: 5-6-08 Date Given to Service: _____

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 - 7110 - 6510 - 6810

Serial #: AR 72 - 0030

Description: A - B - Plus - Screener - Demo

Printer Ser #: AR -

Whole Instrument

Top 1/2

Sim Ser #: _____

Other _____

Probe Ser # DD _____ P -

ACCESSORIES:

☐ 110 V A/C Adapter

☐ Regulator

☐ Mag Card Rdr# _____

☐ Printer Paper

☐ Printer Ribbon

☐ Casio # _____

☐ Mouthpieces

☐ Carrying Case

☐ Dry Gas

Other (Please Specify) _____

Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: _____

Part Number	Description	Qty	Total Cost
<u>MPCAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C W/WHR</u>
<u>B315075</u>	<u>MAIN PCB</u>	<u>1</u>	<u>N/C W/WHR</u>
<u>MPMISC</u>	<u>IR DETECT PCB (6809658) NON STOCK</u>	<u>1</u>	<u>N/C W/WHR</u>
<u>MPCAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C W/WHR</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>2 HR</u>	<u>N/C W/WHR</u>

Repair Notes:

ERROR 71 NOT NOTED - MEMORY WAS @ 23% BUT
NO REAL DATA WAS IN MEMORY.
NO ERRORS, NOTED
CAL W/ QC OPS CHECKS
DURING CAL - JUMPY IR RESULTS AND RANDOM
"LOCK UP" OF UNIT. REPLACE MAIN PCB & IR DETECT PCB.
RE-CAL W/ QC OPS CHECKS

Service Technician: [Signature]

Date: 05-07-08