

Drägersafety

Return & Repair Form

Customer Information:

CUSTOMER Hightown Police Cust # _____
 Date of Return: 11-8-06 Carrier: FEDX Method: 2 DAY

Product Information:

Product: (7110) 7410 6510 6810 Serial # ARTN-0042
 Description: A B PLUS SCREENER _ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Cal ? EC + IR dropping

Part #	Description	Qty	Total
6808455	Fuel Cell	1	
6809512	Cuvette inlet with check valve	1	
MPLABOR	LABOR	1.0	
MPCAL71	7110 CAL	1	

Repair Notes: Replaced questionable FLC and check valve.
Resoldered bad joints on EC motor + switches.
Full cal ops check + O.C.

Technician: QED

Date: 11/8/06