

## Return & Repair Form

### Customer Information:

CUSTOMER West DEPTFORD Police Dep Cust # \_\_\_\_\_  
Date of Return: 6-6-7 Carrier: FED EX Method: Priority

### Product Information:

**Product :** 7110 7410 6510 6810 Serial # ARTN - 0043  
**Description:** A B PLUS SCREENER Printer Serial # \_\_\_\_\_  
**Whole Inst. Top Half** Simulator Serial # \_\_\_\_\_  
**OTHER :** Probe Serial # \_\_\_\_\_

Demo Unit  Returned to stock on: \_\_\_\_\_

### Accessories: (check all that apply):

- 110VAC Adapter
- Regulator
- Mag Card Rdr # \_\_\_\_\_
- Printer Paper
- Printer Ribbon
- Casio # \_\_\_\_\_
- Mouthpieces
- Carrying Case
- Dry Gas
- Other (specify): \_\_\_\_\_

Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: EC Problem

Part #	Description	Qty	Total
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>MP LABOR</u>	<u>LABOR</u>	<u>.5</u>	
<u>MPCALTI</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Replaced leaky F/C Full cal ops  
check + Q.C.

Technician: CRD

Date: 6/7/07