

# Return & Repair Form:

## Customer Information:

Company Name: West Dept B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 4/1/2010 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR TN - 0043  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR - \_\_\_\_\_  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD \_\_\_\_\_ P - \_\_\_\_\_

## ACCESSORIES:

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: FC

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>w/c warr</u>
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>w/c warr</u>
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>w/c warr</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>w/c warr</u>

## Repair Notes:

NEW FC # ARAN-8930  
cal w/QC; ops check

Service Technician: [Signature]

Date: 6/4/2/2010