

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER EAST HANOVER TWP. P.D. Cust # _____
 Date of Return: 5-2-7 Carrier: Fed Ex Method: Ground.

Product Information:

Product: 7110 7410 6510 6810 Serial # ARTN-0046
Description: A B PLUS SCREENER Printer Serial # _____
Whole Inst. Top Half Simulator Serial # _____
OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Repair per attached EC Problem

| Part # | Description | Qty | Total |
|----------------|------------------|-----------|-------|
| <u>0808455</u> | <u>Fuel Cell</u> | <u>1</u> | |
| <u>MPUBOR</u> | <u>LABOR</u> | <u>.5</u> | |
| <u>MPUAL71</u> | <u>7110 CAL</u> | <u>1</u> | |
| | | | |
| | | | |

Repair Notes: Replaced leaky F/c. Full cal ops check + O.C.

Technician: CRD

Date: 5/3/07