

RETURN AND REPAIR FORM



Customer Information

B: _____

S: _____

Company Name: Woodstown Boro P.D., Woodstown, NJ

Date Received: 10/15/13

Date given to service: 10/15/13

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR TN-0051

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: July 2016

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#: _____

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
MP Labor	Labor	1	NC/W

Repair Notes: Replaced defective C156 and C167. Complaint received regarding a problem with units modem receptacle and by changing the two capacitors resolved this issue.

CAL W/QC AND OPS CHECK

Service Technician: BC

Date: 10/17/13