

# RETURN AND REPAIR FORM



## Customer Information

**Company Name:** Woodstown Boro. P.D., Woodstown, NJ

**Date Received:** 03/28/2017

**Date given to service:** 03/28/2017

**Carrier:**  FedEx  UPS  USPS

**Shipping Method:**  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

**Product:**  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

**Serial Number:** AR TN-0051

**Printer Serial#:** AR

**Sim Serial#:** DD

**Probe Serial#:** DD

**Warranty Expires:** \_\_\_\_\_

**Description:**  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

**Test#:** \_\_\_\_\_

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	.5	
6808455	Fuel Cell	1	
6808486	Spacer Plates	2	

**Repair Notes:** Replaced defective fuel cell and spacer plates.

Fuel Cell = ARJL-0033

CAL W/QC AND OPS CHECK

**Service Technician:** BC

**Date:** 04/04/17