

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Riverdale Police Cust # _____
 Date of Return: 2/28/17 Carrier: Fed EX Method: standard

Product Information:

Product : 7110 7410 6510 6810 Serial # ARTN-0056
 Description: A B PLUS SCREENER Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER : _____ Probe Serial # _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter Regulator Mag Card Rdr # _____
- Printer Paper Printer Ribbon Casio # _____
- Mouthpieces Carrying Case Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: ? New Jersey repair EC

Part #	Description	Qty	Total
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>MPLAB02</u>	<u>LABOR</u>	<u>.5</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Replaced leaky F/C. Full cal
ops check + Q.C.

Technician: CRD

Date: 3/1/07