

Return & Repair Form

Customer Information:

CUSTOMER BODOLTON TWP, N.J. Cust # _____

Date of Return: 12/07/2007 Carrier: FX UPS DHL USPS Method: GRD NDA 2DY 3DY
OTHER: _____

Product Information:

Product: 7110 7410 6510 6810 Serial # AR TN - 0061

Description: A B PLUS SCREENER _ Printer Serial # AR -

Whole Inst. Top Half Simulator Serial # _____

OTHER: _____ Probe Serial # DD P -

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter
- Regulator
- Mag Card Rdr # _____
- Printer Paper
- Printer Ribbon
- Casio # _____
- Mouthpieces
- Carrying Case
- Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: CERTICAL

Part #	Description	Qty	Total
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>WARR N/C</u>
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	<u>WARR N/C</u>
<u>WPLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>WARR N/C</u>

Repair Notes: FC IS TH SERIES / REPLACE CELL / CAL W/ QG OPS
CHECKS

Technician: [Signature]

Date: 01-02-08