

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Clementon Police Dept NJ Cust # 54218
 Date of Return: 7/18 Carrier: Fed Exp Method: 2 Day

Product Information:

Product: 7110 7410 6510 Printer Simulator Probe Serial # ARTN-0063
 Description: A B PLUS SCREENER Serial # _____
 Whole Inst. Top Half Printer Serial # _____
 OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: ERROR 41
CAUSED BY GENERATOR? (AT STATION)

W
W
W

Part #	Description	Qty	Total
6810269	MOTOR	1	N/C
8314351	PCB	1	NC
8100902	FLOW SENSOR	1	N/C

Repair Notes: DEPT REP. STATED THAT ERRORS HAPPENED WHEN BACK UP GENERATOR TURNED ON AFTER POWER OUTAGE.
E-41 FLOW SENSOR - REPLACED / E-32 - MOTOR REPLACED - BUT RUNS ALL THE TIME
REPLACED PCB - FULL CAL - QC & OPS ✓

Technician: [Signature]

Date: 07.18.2005