

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER CITATHAN P.I.D Cust # _____
 Date of Return: 4/28 Carrier: FedEx Method: STD

Product Information:

Product: 7110 7410 6510 6810 Serial # ARTN-0065
 Description: A B PLUS SCREENER _____ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter Regulator Mag Card Rdr # _____
- Printer Paper Printer Ribbon Casio # _____
- Mouthpieces Carrying Case Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Problems with SAN-CHANGE and LIN-TEST

Part #	Description	Qty	Total

Repair Notes: Could not reproduce problem. Performed SAN-CHANGE and LIN-TEST without fail.

Technician: CRD

Date: 5/5/06