

Dräger safety

54653

Return & Repair Form

Customer Information:

CUSTOMER Gloucester City Police Cust # _____
 Date of Return: _____ Carrier: FC Method: STD

Product Information:

Product: 7110 7410 6510 6810 Serial # ARTN-0066
 Description: A B PLUS SCREENER _ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Fuel Cell Error Warranty

Part #	Description	Qty	Total
6808455	Fuel Cell	1	N/C
MPCAL71	7110 CAL	1	N/C

Repair Notes: Fuel cell was leaking. Could not see low EC readings. Replaced fuel cell due to leaking. Cal ops check +QC.

Technician: CSD

Date: 8/15/06