

# Return & Repair Form:

## Customer Information:

Company Name: Gloucester City B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 5-13-11 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-~~7110~~-6510-6810-7510 Serial #: ARTN - 0066  
 9510-Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top ½ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return:			
Part Number	Description	Qty	Total Cost
<u>0808455</u>	<u>FC</u>	<u>1</u>	<u>w/cw</u>
<u>0808486</u>	<u>Plate</u>	<u>1</u>	<u>w/cw</u>
<u>mp Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>w/cw</u>
<u>12003</u>	<u>Hose Barb</u>	<u>1</u>	<u>w/cw</u>
<u>mp Labor</u>	<u>Labor</u>	<u>.5</u>	<u>w/cw</u>

Repair Notes:

FCTT ARCA-0826  
Replace broken Hose Barb, Hose Barb  
was broken upon arrival.  
Cal w/ DC & OPS check

Service Technician:



Date: 5-16-2011