

**Return & Repair Form:**

**Customer Information:**

Company Name: Millville P.P. B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 8-11-89 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY Other

**Product Information:**

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR TN - 0071  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

**Repair Information:**

Test #

Reason for Return: FC

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>n/c warr</u>
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>n/c warr</u>
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>n/c warr</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>n/c warr</u>

Repair Notes:

Replace FC  
Cal w/ QC & Ops Check

Service Technician: [Signature]

Date: 08-13-09